


ACL reconstruction

Adjustable juxtacortical fixation

PULLUP[®]

- ▶ Strength and rigidity
- ▶ Single adjustable size
- ▶ Fast and simple technique



Pull, flip, adjust.

PULLUP®

Solution for fixation of knee ligament reconstructions:

Reproducible


Controlled locking at blind tunnel endstop, conservation of bone stock

Simple

No pre-op planning of tunnel length, perioperative adjustment

Reliable


Stable locking system, renowned materials



A system that can be adapted to your surgical approach


The PULLUP® system is available in two versions:

PULLUP®



For cortical tunnels with \varnothing 4.5 mm


PULLUP® XL



For cortical tunnels with \varnothing from 5 to 10 mm

Surgical technique

Pullup® fixation characteristics:




1



Implant-graft fixation

Harvest and clean the graft (semitendinosus/gracilis or short 4-strand hamstring graft). Select either a PULLUP® or PULLUP® XL implant depending on the diameter of the cortical tunnel: PULLUP® for a diameter of 4,5 mm and PULLUP® XL for a diameter of 5 to 10 mm. Remove the implant from its protective mount sheet by simultaneously pulling on the two white loops. Insert the graft through the two implant loops, then fold back and suture the graft.


2



Drilling the femoral tunnel

Position an Ø 2.5 mm eyeletpin and push it until it appears beneath the surface of the skin. Perform an incision. Drill with the Ø 4.5 mm cannulated drill bit until the external cortex is penetrated through (for PULLUP®). Next, drill a blind tunnel to the desired length (15 mm minimum) corresponding to the graft diameter.

3




Preparing the graft placement

Use an eyeletguidepin to introduce a suture passer through the tunnel.


Pass all the implant threads through the suture passer. Tension all the threads by using the suture passer and use the white/blue thread to pull the button through the external cortical surface.

4



Flipping the implant


Flip the plate by pulling the green thread.



Locking the fixation

Pull alternately on the white threads until the implant is locked into place.


Pull on the graft to check for stability.



Removing the threads

Cut the white threads with the cutting pliers.

Remove the green and black threads by pulling on either end.



Tibial fixation

Apply a tension by pulling alternatively on the white threads of the PULLUP XL® implant until it is locked, cycle the knee in flexion / extension, and then tension the system once more.



Removing the threads

Control the lack of resistance on the flexion of the knee, then release the green and white/blue threads by pulling one of their extremities. Use the cutting pliers to cut the white threads of the implant.

Ordering information

PULLUP® adjustable juxtacortical fixation system for ligament reconstruction

Codes	Designation	Packaging
PULLU01201	PULLUP® adjustable juxtacortical fixation system in cortical tunnel= ø 4.5 mm	1
PULXL01202	PULLUP® XL adjustable juxtacortical fixation system in cortical tunnel diameter 5 to 10 mm	1

Speak to your local Business Development Manager for further information or contact us using the details below:

T: 01443 719 555

E: customerservice@hiuk.co.uk

www.hospitalinnovations.com

Hospital Innovations Limited

Concept House

Talbot Green Business Park

Pontyclun

CF72 9FG

